











CoCo-80X Menu Guide

Button Name	Functions
 Power	Power on/off the CoCo-80X or lock keypad
 Save	Save the signals designated by the CSA project
 Setup	Go to the main setup page
 Analysis	Open the "Analysis Groups" page to choose a CSA project
 Display	Switch to the current CSA project signal display (must have an active project running)
 File	View existing record files stored on CoCo memory.
 Input	Enter the input channel setup page
 Trace Selection	Go to previous or next trace window
 Recording	Start/Stop recording time stream(s)
 	Save signal(s)

From:

<https://help.go-ci.com/> - **Crystal Instruments Help**

Permanent link:

<https://help.go-ci.com/coco:menu>

Last update: **2023/08/29 21:20**