



# Setup a Post Analyzer test

Before starting, a file with recorded data in the form of .atfx, .csv, or .txt should be accessible to add into EDM.



New Project Wizard ✕

**Enter project information and select data source file**

**Project information**

Project name

Project description   Create project by using a template

**Templates**

Select	Project name	Description

**Select data source file**


**Tips**

Press [Add single source file] button to add one or more files with the same structure. PA can analyze them one by one.

Add single source file (Browse and drag files from [Data File Browser](#))

< Back Next > Cancel

New Project Wizard

 **Enter project information and select data source file**

**Project information**

Project name:



Project description:

Create project by using a template

**Templates**

Select	Project name	Description


**Select data source file**

- REC0120\_PA1.atfx 
  - Block(Ch1), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
  - Block(Ch2), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
  - Block(drive), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
- SIG0010.atfx 
  - Block(Ch1), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
  - Block(Ch2), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
  - Block(Ch3), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
  - Block(drive), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)

**Tips**

Press [Add single source file] button to add one or more files with the same structure. PA can analyze them one by one.

(Browse and drag files from [Data File Browser](#))



In the beginning, you will see the New Project Wizard window.

1. Choose your project type, and click **Next**.
2. Insert your project name, and click **Add single source file** to add your source file
3. Click the checkboxes under the file just added to be included in the analyzation, and click Next
4. Choose the signal types to be computed

From:  
<https://help.go-ci.com/> - **Crystal Instruments Help**

Permanent link:  
<https://help.go-ci.com/pa:setup>

Last update: **2025/06/24 21:00**