

Setup a Post Analyzer test


Before starting, a file with recorded data in the form of .atfx, .csv, or .txt should be accessible to add into EDM.


In the beginning, you will see the New Project Wizard window.

1. Choose your project type, and click **Next**.
2. Insert your project name, and click **Add single source file** to add your source file
3. Click the checkboxes under the file just added to be included in the analyzation, and click Next
4. Choose the signal types to be computed




New Project Wizard
Enter project information and select data source file
Project information
Project name
Spectral
Project description
Create project by using a template
Templates
Select
Project name
Description
Select data source file
Tips
Press [Add single source file] button to add one or more files with the same structure. PA can analyze them one by one.
Add single source file
(Browse and drag files from Data File Browser)
< Back
Next >
Cancel


New Project Wizard
Enter project information and select data source file
Project information
Project name
Spectral
Project description
Create project by using a template
Templates
Select
Project name
Description
Select data source file
REC0120_PA1.atfx
Block(Ch1), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
Block(Ch2), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
Block(drive), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
SIG0010.atfx
Block(Ch1), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
Block(Ch2), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
Block(Ch3), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
Block(drive), Block size = 1024 Sampling rate =5.12 kHz Duration =0.2 (s)
Tips
Press [Add single source file] button to add one or more files with the same structure. PA can analyze them one by one.
Add single source file
(Browse and drag files from Data File Browser)
< Back
Next >
Cancel



From:
https://help.go-ci.com/ - **Crystal Instruments Help**

Permanent link:
https://help.go-ci.com/pa:setup?rev=1750798452

Last update: **2025/06/24 20:54**