2025/11/01 03:33 1/2 VCS Random Testing

# **VCS Random Testing**

## **Create New Test**

# **Test Configuration**

# **Input Channels**

## **Channel Type**

**Control** vs. **Monitor**. When running a test, there will need to be at least one control sensor. The control sensor is used to monitor the actual vibration levels that the shaker is producing. It then sends this data to the controller so that it maintains the targeted profile. This sensor should be mounted somewhere on the shaker/slip table itself, not the Device Under Test (DUT). Monitor sensors will show the levels that the DUT itself is experiencing.

## Measurement Quantity

Select the type of quantity that the sensor is reading. For accelerometers this will be acceleration.

## **Sensitivity**

This is the output sensitivity of the sensor. This value states how many mV the sensor outputs per engineering unit of measurement. In most accelerometer cases this value will be in mV/g and can be found with the sensor.

#### Input Mode

There are five modes in which the inputs can operate.

*DC-Differential-* In the DC-Differential mode, neither of the input connections is referenced to the local ground. The input is taken as the potential difference between the two input terminals, and any potential in common with both terminals is canceled out. The Common Mode Voltage (CMV) will be rejected as long as the overall input voltage level does not saturate the input gain stage. Beware that very high CMV will cause clipping and may damage the input circuitry. Signals with a nonzero mean (DC component) can be measured in this mode.

*DC-Single End-* In single-ended mode, one of the input terminals is grounded and the input is taken as the potential difference of the center terminal with respect to this ground. Use this mode when the input needs to be grounded to reduce EMI noise or static buildup. Do not use this mode when the signal source is ground referenced or ground loop interference may result. This mode also allows signals with a non-zero mean to be measured.

AC-Differential- AC-Differential is a differential input mode that applies a low-frequency high-pass (DC-blocking) analog filter to the input. It rejects common mode signals and DC components in the input

signal. Use this when DC and low-frequency AC voltage measurements are not required or when a DC bias voltage is present. The analog high-pass filter has a cutoff frequency of -3dB at 0.3 Hz, and -0.1dB at 0.7 Hz for the IEPE input mode.

AC-Single End- AC-Single End grounds one of the input terminals and enables the DC-blocking analog filter. Use this mode for non-ground referenced sources where measuring the DC or low-frequency components are not required. It shares the same high-pass filter as that of AC-Differential.

*IEPE (ICP)*- All Crystal Instruments products support IEPE (Integral Electronic PiezoElectric) constant current output type input channels. IEPE refers to a class of transducers that are packaged with built-in voltage amplifiers powered by a constant current. These circuits are powered by a 4 mA constant current source at roughly 21 Volts.

# **Running the Test**

Now that all of the parameters for the test have been setup, the test is ready to run. Here are the final steps to start the test.

- 1. Press the **Connect** button to connect to the controller.
- 2. Press the Run button.
- 3. The **Spider Check List** will now appear. Here you can check the settings of the test and verify that all is set up properly. Press **Start** once this has been verified.
- 4. The Pre-Test will now begin. This will verify that the control loop is properly established and provide data that the control loop needs.
- 5. The test is now running

From:

https://help.go-ci.com/ - Crystal Instruments Help

Permanent link:

https://help.go-ci.com/vcs:random?rev=1714491609

Last update: 2024/04/30 15:40

https://help.go-ci.com/ Printed on 2025/11/01 03:33