Vibration Control System (VCS)

Getting Started

Refer to the links below to get started on setup:

1. Important Safety Information
2. Connecting Your PC to the Spider Front-End
3. Running EDM
4. Where is My Data?

Training

Check out our guides to set up some of the most frequently used tests and learn some of the features and capabilities that VCS software has to offer.

Random
Random excitation is often used to simulate real-world vibration. The purpose of the random vibration control system is to generate a true random drive signal such that, when the signal is applied via an amplifier/shaker to the device under test, the resulting shaker output spectrum will match the user-specified test profile. VCS also has the capability to perform Sine on Random (SoR) and Random on Random (RoR) testing.

Swept Sine
Unlike Random testing which generates many frequencies at once over the band of interest, Swept Sine testing generates energy at one frequency and sweeps this frequency through a preset range.

Classic Shock
A Shock test outputs a series of pulses to excite the structure under test. The response is measured at one or more locations on the structure and a spectral analysis is used to determine its response and resonance characteristics. This pulse response approximates the impulse response, the theoretical response to an infinitely tall and narrow spike function with an area of unity.

FAQs

Take a look at some of our helpful guides to some of the most frequently asked questions.
Don't see what you're looking for below? Send us an email at **support@go-ci.com**.

Importing a Profile into VCS

Setting up Limit Channels

How to Record Timestream Data During Testing

How to Change Engineering Units

Report Templates

Test Types

Controller Disconnects During Long Tests - Check Windows Power Management

From:
https://help.go-ci.com/ - **Crystal Instruments Help**

Permanent link:
https://help.go-ci.com/vcs?rev=1793129133

Last update: **2026/10/27 19:25**